

Permission to print:	Yes
Near Miss or Accident	Accident
Type of incident:	Management
Catagory	Circuit disruption
Knowledge Error	No
Protocol issue	No
Rule Error	No
Skill Error	No
Team Issue	No
Violation	No
Chance event:	Yes
Description:	While Perfusing for a AVR + CABG case, venous cannulae came out of right atrium so had no venous drainage while venous line was filled again..
Contributing factors:	Nil known
Corrective action:	Stopped the pump refilled venous line by swapping the prime line to venous line and continued bypass..It took roughly 1.5 mins start bypass again..
Preventative action plan:	n/a
Manufacturer advised:	No
Discussed with team:	Yes
Ext Authority Advised	No
Patient outcome variance f	Nil