



Australian and New Zealand College of Perfusionists

PIRS 2 ALERT

Device: Inspire 6M Oxygenator

PIRS2 has received 2 reports from one centre 4 days apart of a subtle fluid leak across the arterial temperature port during priming. PIRS advise a careful check for this while awaiting further details.

The common lot number was 2001160035.



Please contact pirs@anzcp.org for further information.